

CERTIFICATE

Biocompatibility Test

Material tested:

Vitallium

Dental alloy for removable restorations

Composition, w/w%:

Co 60.0; Cr 31.5; Mo 6.0; Si, Mn, C < 1 %

Manufacturer:

Dentsply International Prosthetics Division
570 West College Avenue
York, PA 17405-0872
USA

EU Distributor:

Elephant Dental B.V.
Verlengde Lageweg 10
1628 PM Hoorn
The Netherlands

The dental alloy Vitallium was tested by the Rheinisch-Westfälischen Technischen Hochschule (RWTH) Aachen on:

Cytotoxicity:

The cytotoxic potential of the dental alloy was tested in vitro with L-929 fibroblasts.
Method: „Test on extracts“ and in direct contact.

Test result:

Vitallium had no cytotoxic potential.

Hoorn, 11.01.2005

T.J. Grinwis
QA Director

